



GURU GOBIND SINGH MEDICAL COLLEGE, FARIDKOT. (PUNJAB)- 151203
(Constituent Medical College of Baba Farid University of Health Sciences, Faridkot)

E-mail: procurement@ggsmch.org Website: www.ggsmch.org

No. Purchase/GGS/2023/ 8532

Date: 20/11/2023

.....
.....
Sub: Quotation for Supply and installation of VDRL Shaker required at the department of Microbiology, GGS Medical College & Hospital, Faridkot.

Sealed quotations are invited for supply and installation of **VDRL Shaker** on terms & conditions given as under.

Sr. No.	Name of the Item	Specifications	Make/Brand/Guarantee/warranty	Required Quantity	Amount (in Rs.) including installation
1.	VDRL Shaker	Attached herewith	To be filled by the bidder	01	

Terms & Conditions:

1. The material should be good quality and according to the requirement.
2. The material should meet standards in Quality and as per required. Make/ Brand and Specifications should be mentioned clearly.
3. Supply should be F.O.R. destination at Store G.G.S. Medical College & Hospital, Faridkot.
4. Rates quoted should not be more than those quoted to DGS&D and any other central or State Govt. Organizations.
5. Payment will be made after getting satisfactory report from the concerned department.
6. Supply of the item is to be made within 30 days from the date of issue of the supply order. If the supply is not made within the stipulated periods then late delivery charges @2% will be imposed on the total amount up to delay of 30 days and thereafter @4% for another 30 days. In case the goods are not supplied, the competent authority will take action as deemed fit.
7. Taxes should be clearly mentioned separately.
8. Guarantee/Warranty should be clearly mentioned.
9. Validity of Rates:- 90 days from the last date of receipt of Quotations.

Note: Only Terms & Conditions mentioned on this Quotation will be considered for supply order.

You are requested to send your lowest bid in sealed envelope, addressed to The PRINCIPAL, G.G.S Medical College, FARIDKOT super scribing "QUOTATION" for "VDRL Shaker and Quotation no..... date....." on the top of the Envelope.


Last Date for receipt of Quotation /Tender in Principal Office is 08/12/2023 by 5.00p.m. through Registered/ Speed Post/Trackable Courier Only.

Principal

Specifications of VDRL Shaker

Parameter	Description
Platforms size	12 inch *12 inch,
Outer constriction	SS 304 or Mild steel with powder coated
Speed range	50-300 RPM
RPM accuracy	+/- 5 RPM
Power	220/230 V, AC supply, power cord
Display	LED
Timer	Automatic
Use	For VDRL plates, slide, Agglutination
Motor	Heavy duty motor
Certification	ISO,CE Certificate must be provided

- ਇਹ ਤਸਦੀਕ ਕੀਤਾ ਜਾਂਦਾ ਹੈ ਕਿ ਉਪਰੋਕਤ Specification tailor made ਨਹੀਂ ਹਨ।


Dr. Shilpa Arora
Associate Professor
Microbiology Deptt
GGSMC, Faridkot


Dr. Shipra Galhotra
Assistant Professor
Microbiology Deptt
GGSMC, Faridkot


Dr. Nitika Dhuria
Assistant Professor
Microbiology Deptt
GGSMC, Faridkot



