

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital, Faridkot.

E-Tender Form

(E-Tender enquiry for Supply and Installation of equipments required for GGS Medical College & Hospital, Faridkot.

Tender Notification No :	To be provided by the E-procurement portal of the Govt. of Punjab.
Tender Notification Date:	13/01/2021
Requirement	E-Tender notice for supply and Installation of equipments as per LIST
Cost of the tender document:-	Rs.2000/- (Non-refundable) to be deposited through Online Mode Only in favor of Registrar, Baba Farid University of Health Sciences, Faridkot.
Tender Processing Fee	To be charged by Govt. of Punjab as per its norms. (Non- refundable)
Earnest Money Deposit (EMD)	As per List. The Earnest Money Deposit must be submitted in the shape of Online Payment in favor of Registrar, Baba Farid University of Health Sciences, Faridkot on or before due date (Refundable to the Non-successful bidders, without any type of interest or other charges). However, it will be converted into Performance security in case of successful tenderer and will be returned after successful completion of the contract period.
Date of start of downloading of tender documents	13.01.2021 from the website of the Punjab Government i.e. https://eproc.punjab.gov.in
Website for downloading of the tender document:-	https://eproc.punjab.gov.in However, the details may also be obtained from the University website i.e. www.bfuhs.ac.in and college website www.ggsmch.org
Last date for downloading of the tender document:-	10.02.2021 up to 12.30 pm
Last date & time for uploading of the tender documents:-	10.02.2021 up to 1.30 pm (through online mode only)
Date, time and venue for opening of the Technical Bids	By the next day from the last date of submission of tenders (by 5:00 p.m.) on the e- procurement portal of the Govt. of Punjab in , Baba Farid University of Health Sciences, Faridkot
Date, time and venue for opening of the Price Bids	17.02.2021 on the e- procurement portal of the Govt. of Punjab in, Baba Farid University of Health Sciences, Faridkot
Who can be contacted for obtaining more information about the tender.	Principal, Guru Gobind Singh Medical College & Hospital, Sadiq Road, Faridkot. 01639-251111, 98773-65600, 88472-19897 E-mail: ggsmc@punjab.gov.in , pr_ggsmc@yahoo.com (on all working days from 9.00 a.m. to 5.00 p.m.)

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital, Faridkot.

NOTICE INVITING E-TENDER

E-Tenders are invited on or before **10.02.2021** from manufacturers or their authorized agents/distributors **for supply and Installation of equipments required at GGS Medical College & Hospital, Faridkot.** The tender document containing detailed terms & conditions may be downloaded from the E-procurement website of the Punjab Government i.e. <https://eproc.punjab.gov.in> and its detail may also be seen at the University website www.bfuhs.ac.in and college website www.ggsmch.org

TERMS AND CONDITIONS:-

1. The tender must be uploaded on or before the last date/ time of the submission of tender.
2. The Tender processing fee should be submitted through Net Banking/Credit card/Online mode only and as per Punjab Govt.
3. The **tenders will be opened online** By the next day from the last date of submission of tenders (by 5:00 p.m.) on the e- procurement portal of the Govt. of Punjab in , Baba Farid University of Health Sciences, Faridkot on the website i.e. <https://eproc.punjab.gov.in> at the, Baba Farid University of Health Sciences, Faridkot. The bidder(s) shall be at liberty to be present, in person or through their authorized representative(s) at the time of opening of the tender as specified in the Tender Notice. In case the authorized representatives are to be present, they must furnish the authority letter from the bidder (s), on whose behalf they are representing otherwise they will not be allowed to participate in the process of opening of tender.
4. **The Price bids of technically qualified bidders will be opened** on **17.02.2021** at 2.30 p.m. on the website i.e. <https://eproc.punjab.gov.in> at, Baba Farid University of Health Sciences, Faridkot. In case of any change of date and time it will be notified to the technically qualified bidders through E-mail/telephone.
5. The Registrar/Principal reserves all rights to accept or reject any or all the tenders without assigning any reason.

Registrar

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital, Faridkot.

INSTRUCTIONS/ GUIDELINES TO THE TENDERERS

1. The bidder needs to register himself/ herself on <https://eproc.punjab.gov.in>. The bidder is also required to obtain Class III digital signature certificates to complete this process.
2. Please download the Tender document from the website of e-procurement of the Govt. of Punjab <https://eproc.punjab.gov.in>. Please fill all the relevant blanks on all the pages of the tender document sign along with a stamp/ seal all pages and then a scanned copy of the same may be uploaded on the website at the time of submission of the tender document.
3. **It should be clearly noted that this tender will be accepted though e-tender mode only.** The tenders submitted through offline mode will not be accepted under any circumstances.
4. **Tender Fee** of Rs.2000/- (non-refundable) may be deposited through online Mode Only.
5. **Tender Processing Fee:** as per Punjab Govt. (non-refundable) may be deposited through online mode i.e. Net Banking/ Credit Card/ Debit Card only. The tender processing fee will not be accepted through any other mode. .
6. **Refundable Earnest Money Deposit (EMD): As per list** to be deposited online mode only in favor of Registrar Baba Farid University of Health Sciences, Faridkot
7. **Upload** signed copy of Technical Bid Compliance Statement (**Annexure-I**).
8. **Upload** an affidavit regarding Non-Black listing as per proforma given at **Annexure-II** duly attested by an Executive Magistrate or a Notary Public.
9. In case the Bidder is Authorized Supplier/Agency, the Authorization Certificate as per the Format given at **Annexure-‘III’** (duly filled in), **to be uploaded.**
10. In case the Bidder is Authorized Supplier/Agency, an undertaking/certificate issued by their Principal Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply of the Instruments then the Principal Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC of the Equipments/Instruments. (**Annexure – ‘IV’**), **to be uploaded.**
11. **Upload** details of Bank Account for refund of EMD (**Annexure – V**).
12. In addition to this, following **documents are to be uploaded** with Technical Bid:-
 - i) Details of registration as Company /Firm/ Establishment.
 - ii) Certificate regarding **standard in quality EU-CE /US FDA or any other** as per in required specifications.
 - iii) Copy of Certificate of Registration for service Tax/TIN/TAN/PAN/VAT.
 - iv) A certificate from C.A. regarding required Turnover with Balance Sheet for the last 3 (three) financial years i.e. 2017-18, 2018-19 and 2019-20.
 - v) Copy of the IT Returns for three financial years i.e. 2017-18, 2018-19 and 2019-20.
13. Price should be quoted and **uploaded** only in Excel Sheet proforma at **Annexure-‘VI’**.

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.***

LIST OF EQUIPMENTS

Sr.	Name of Equipment	Approx. Qty. Reqd.	Earnest Money to be Submitted (in Rs.)
1.	OT Table (Electrohydarulic)	05	1,00,000/-
2.	Electro Cauty (Electrosurgical Unit)	05	20,000/-

- The quantity may increase /decrease as per requirement.
- The bidder must mention regarding tentative delivery time of material.
- The bidder must quote AMC and CAMC in percentage on Company / Firm's letter head of quoted items.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital, Faridkot.

Detail of Items with specifications:

OT Table Electrohydarulic

The operation table should have the following features:-1) Four/Five Section Table top with divided Foot section. Table top should permit x-ray penetration, good quality durable composite material and fluoroscopy with full length X-ray cassette tunnel accessible from either end. There should be U cut compatible with gynecological surgery.

Should have a remote handset for control of various functions and manual override panel. There should be additional inbuilt control enabling full use of table in case of handset failure. Should return to level function with one button press.

All table positioning, i.e., height, back section, lateral tilt, trendelenburg, and anti-trendelenburg, should be operated electro-hydraulically through handset auxiliary control on table body

The casings on the frame and centre supporting column should be made of hygienic stainless steel. Mattress of minimum 50 mm thickness high quality and latex free and should be radiolucent and suitable for fluoroscopy.

Table should have mobile base with lockable castors. Should have built in kidney bridge.

Measurements (Approx): Minimum Height 650-690mm, Maximum Height: 1040-1100mm , Side tilt: +/-20degrees , Back section adjustment: - 40 degrees to 80 degrees , Foot section adjustment: - 90 to 0 degree, detachable , Trendelenburg: 30 degree , Anti trendelenburg: 30 degree Head section adjustment: +/- 45 degree, detachable Width: 550 mm or better, Length: 2000 mm or better , Cranial and caudal traversing 200 300 mm , Maximum patient weight 300 Kg, Width: 550 mm or better, Length: 2000 mm or better.

Certificate regarding standard in quality: - US FDA / European CE / Bureau of Indian Standards (BIS)

Warranty / Guarantee: - 05 years

The table should be supplied with the following accessories:-

- Padded arm rest with straps - pair with clamps
- Anesthesia screen with clamps
- Side supports: pair with clamps
- Shoulder supports: pair with clamps
- Knee crutches: pair with clamps
- X-ray cassette tray with pushing rod
- Accessories for operating in prone position, with lithotomy poles and attachments.

Price of above accessories should be quoted too separately in price bid. The specific accessories and their quantity will be decided on the basis of actual requirement.

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital, Faridkot.

Advance Electro-surgical generator system

Specifications

- ❖ Microprocessor / microcontroller technology for monopolar and Bipolar surgeries
- ❖ Should Automatically monitor tissue impedance & adjust power output to reduce damage & drag. featuring Constant Control Technology & contact quality monitoring for return electrode.
- ❖ Should have a fully automated Preset advanced cutting (Low & High) cut mode to provide a scalpel like cut effect with minimal thermal necrosis
- ❖ Should facilitate skin incision with minimal to no scarring, facilitates plasma cut.
- ❖ Displays should be Large & readable from a distance
- ❖ Different audible tones to differentiate from Cut, Coag and Bipolar modes when activated
- ❖ Active mode setting should illuminate when in use for visual confirmation
- ❖ Should give ERROR codes in case of any Errors & fault in the unit
- ❖ Should have an easy-to-use 'reset' function
- ❖ Should have 2 receptacles each of Hand & Foot control for Monopolar output
- ❖ Should be activated either by Foot pedal or Hand control at surgeon's discretion without any switches or change over in the equipment to do so from any of the receptacles.
- ❖ Should have Soft Coagulation for slower thermal penetration.
- ❖ Should have Auto Bipolar to provides automatic activation when the electrode touches tissue, combined with automatic stop based on the impedance of the tissue
- ❖ Should have an in-built current meter that shows the flow of current in Bipolar mode & indicate tissue dehydration.
 - ❖ Should be defibrillator proof generator & should be denominated on equipment.
 - ❖ Should be US FDA approved under section 510, ISO, IEC certification.
 - ❖ Should meet all applicable IEC standards
 - ❖ Should be compatible with Argon Plasma & Smoke Evacuators
 - ❖ High Frequency RF Leakage should be less than 150m A (<150 m A)
 - ❖ Should Have RS232 & USB communication Ports.
 - ❖ Should have PER >97% for effective output in surgeries.
 - ❖ CREST Factor should be > 8 in Spray Mode for effective tissue effect.

Technical specification:

- ❖ E.S. Generators operating frequency: 400KHz
- ❖ Equipment classification: IEC 60601-1 Class 1
- ❖ Equipment type: IEC 60601-1 Type CF
- ❖ Monopolar should have a minimum 3 cut modes
 - 1) An advanced low cut & high cut Mode, Geometric electron modulation technology to give scalpel like effect- with constant voltage, which enables surgeon to cut through the skin with minimal damage to tissue.
 - 2) Pure Cut Mode - Power up to 300 watts
 - 3) Blend Mode- Power up to 200 watts

Baba Farid University of Health Sciences, Faridkot

E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital, Faridkot.

-
-
- ❖ Monopolar should have a minimum 4 Coagulation modes including Soft Coagulation up to 120W each Coag mode
 - ❖ Should have 3 Bipolar modes Micro, Macro & Auto Bipolar up to 80 watts each mode

Should come with following accessories:

Capital

- 1- Electrosurgical Generator
- 2- Monopolar Foot Switch
- 3- Bipolar Foot Switch
- 4- Universal Adapter

Open Surgery

- 1) Disposable nonstick cautery pencil for Open Surgeries - 50 nos
- 2) Disposable return electrode with cord- 50 nos

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.***

TERMS AND CONDITIONS

**For Supply and Installation of equipments required for GGS Medical College & Hospital,
Faridkot**

ELIGIBILITY

- The sole manufacturers of equipments or their authorized agents/distributors may quote their rates.
 - In case of Authorized Supplier/Agency/Distributor, the Authorization Certificate as per the Format given at **Annexure-‘III’** should be uploaded.
 - In case the Tenderer is authorized dealer/supplier an undertaking/certificate issued by their Principal Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply then the Principal Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC. (**Annexure – ‘IV’**).
1. This institution reserves the right to reject tenders without assigning any reason and increase or decrease the quantity of the articles tendered.
 2. If the supply is not made within the stipulated period then late delivery charges @**2%** will be imposed on the total amount of Supply Order up to delay of **30 days** and thereafter @ **4%** for another **30 days** after which Supply Order will be deemed cancelled & security/earnest money forfeited and company will be black-listed for future.
 3. In-complete or conditional offers incorporating price variation will not be entertained.
 4. The firm should have been in existence for at- least **three years** and it should have turnover of minimum **Rs.1,00,00,000/- per year**.
 5. The successful bidder shall deposit bank guarantee of 10% of the value of the equipment in the shape of Demand Draft only.

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.***

Annexure-I

TECHNICAL BID COMPLIANCE STATEMENT

Name and Address of the applicant / firm _____

Specify whether Manufacturer/Dealer/Distributor: _____

Sr. No.	Particulars	Remarks
1.	Tender Fee of Rs.2000/- through Online Mode only on or before due date.	Yes/No
2.	Tender Processing fee charged by Punjab Govt. as per their norms. (Non-refundable).	Yes/No
3.	Earnest Money as per LIST in favor of Registrar Baba Farid University of Health Sciences, Faridkot through online mode only.	Yes/No
4.	Technical Bid Compliance Proforma uploaded (Annexure-I).	Yes/No
5.	Whether an affidavit regarding Non-Black listing as per proforma given at Annexure-II duly attested by an Executive Magistrate or a Notary Public uploaded.	Yes/No
6.	In case the bidder is Authorized Supplier/Agency, the Authorization Certificate as per the Format given at Annexure-‘III’ uploaded.	Yes/No
7.	In case the Tenderer is Authorized Supplier/Agency, an undertaking/certificate issued by their Principle Manufacturer/Supplier that in case dealership/distributorship is withdrawn after supply then the Principle Manufacturer/Supplier will be responsible for Guarantee/Warranty/AMC/CMC (Annexure – ‘IV’) uploaded.	Yes/No
8.	Details of Bank Account for refund of EMD (Annexure – V) uploaded.	Yes/No
9.	Price Bid in the prescribed format in Excel Sheet (Annex – VI) uploaded.	Yes/No
10.	Copy of Certificate of Registration for service Tax/TIN/TAN/PAN uploaded.	Yes/No
11.	A certificate from C.A. regarding required Turnover with Balance Sheet for the last 3 (three) financial years i.e. 2017-18, 2018-19 and 2019-20 uploaded.	Yes/No
12.	Copy of the IT Returns for three financial i.e. 2017-18, 2018-19 and 2019-20 uploaded	Yes/No
13.	Certificate regarding standard in quality.	Yes/No
14.	E-mail ID	Yes/No

Signature & seal of bidder

Place:

Date :

Note: Please upload Catalogue/Brochure/Pamphlets with complete specifications of each Equipment quoted.

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.***

Annexure-II

(To be furnished on non-judicial stamp paper worth Rs.30/- duly attested by Executive Magistrate or Notary Public).

AFFIDAVIT

I/We _____
partner/sole proprietor (Strike out which is not applicable) of (Name & Address of Firm)
_____ do hereby declare and solemnly affirm:-

- a) That the individual/firm/ companies are **not debarred or black- listed** by any department of Union/ State Government or any autonomous institute.
- b) That no partner or shareholder, directly or indirectly connected with the applicant who has been debarred or blacklisted by any department of Union Govt./State Govt./Autonomous Institute.
- c) And that the terms and conditions for supply and Installation of Equipments at GGSMCH, Faridkot, are acceptable to me/us. I/We shall abide by them in letter and spirit.

Date:

Place:

DEPONENT

VERIFICATION

I/We do hereby solemnly declare and affirm that the above declarations are true and correct to the best of my/our knowledge and beliefs. No part of it is false and nothing has been concealed therein.

Date:

Place:

DEPONENT

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.***

Annexure- III

MANUFACTURER'S/PRINCIPLE'S AUTHORIZATION FORM

TO

The Registrar
Baba Farid University of Health Sciences,
Faridkot -151203

Ref. No.....

Dated:

**Sub: Authorization Certificate in favour of M/s..... for supply
of**

We, M/s....., who are established and reputable
manufacturers of(name of equipment) having factory(ies) at
..... and, hereby authorize
M/s.....(name and address) to bid, negotiate and conclude
the Tender formalities with you against Tender No..... for the above
equipment(s) manufactured by us.

No company or firm or individual other than M/s..... are authorized
to bid, negotiate and conclude the tender formalities in regard to this business against this
specific tender.

We, hereby extend our full guarantee and warranty as per the conditions of tender for
the goods offered for supply against this tender by the above firm.

Yours faithfully,

(Name)

For and on behalf of M/s _____
(name of manufacturer/Principle)

**Note: This letter should be signed by a person competent and having authority to sign
on behalf of manufacturer, and should be on manufacturer Letter Head and same will
be uploaded with Technical Bid.**

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.***

Annexure - IV

UNDERTAKING BY MANUFACTURER/PRINCIPLE SUPPLIER

TO

The Registrar
Baba Farid University of Health Sciences,
Faridkot -151203

Ref. No.....

Dated:

Sub: Undertaking for after sales service for (Name of Equipment)

We, M/s....., who are established and reputable manufacturers of(name of equipment) have authorized M/s.....(name and address) to bid, negotiate and conclude the Tender formalities with you against Tender No..... for the above equipment(s).

Further, we undertake that in case dealership/distributorship is withdrawn after supply of equipment then we shall be responsible for after sales service till the date of guarantee/warranty of the equipment and afterwards for a period of 10 years.

Yours faithfully,

(Name)

For and on behalf of M/s_____ (name of manufacturer/Principle)

Baba Farid University of Health Sciences, Faridkot

***E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.***

Annexure- V

Details of Bank Account of the firm who has deposited EMD

Name of the firm: _____

Sr. No.	Particulars	Detail
1.	Account No.	
2.	Name of Bank	
3.	Branch Name	
4.	IFSC Code of Bank	
5.	Name of Operator	

Baba Farid University of Health Sciences, Faridkot

*E-TENDER NOTICE FOR procurement of equipments for GGS Medical College & Hospital,
Faridkot.*

ANNEXURE - VI

PRICE BID

TO BE UPLOADED in Printed/Computerized format in Excel Sheet Attached.

Hard Copy of Technical Bid Should also be submitted on or Before Due date in a sealed envelope.