



**GURU GOBIND SINGH MEDICAL COLLEGE, FARIDKOT. (PUNJAB)- 151203.**

(Constituent Medical College of Baba Farid University of Health Sciences, Faridkot)

E-mail: [pr\\_ggsmc@yahoo.com](mailto:pr_ggsmc@yahoo.com) Website: [www.ggsmch.org](http://www.ggsmch.org)

Phone: 01639-251111 Fax : 01639-251070

No.Purchase/GGS/2015/..17550....

Dated...01.07.2015

**Sub: Quotation for supply and Installtion of Equipments required at this Institution.**

Sealed quotations are invited for supply and Installation of Equipments on terms & Conditions given as under.

Sr No.	Name of Item/Equipment	Specifications
1.	Multipara Monitor	<ul style="list-style-type: none"><li>• 10" Colour TFT LCD display with</li><li>• waveforms upto 8 Channels</li><li>• Portable , streamline handle design</li><li>• Monitor measures from Adult to Neonates</li><li>• easurement of one channel ECG anytime, anywhere you like</li><li>• Display of ECG waveform, heart rate, analysis results and battery status on the backlit LCD</li><li>• Measurement results transmission to PC by USB mode</li><li>• Auto power-off while no key is pressed in 30 seconds</li><li>• Auto-evaluation based on detected heart data</li><li>• Dual measuring modes</li><li>• 2 batteries can take at least 500 times measurement</li><li>• Data Review freely</li></ul>
2.	Portable Suction machine	<ul style="list-style-type: none"><li>• Portable suction unit: size 38x17x28.5cm, weight 5.1kg</li><li>• Description</li><li>• Technical specifications: SUCTION MACHINE with pump: -</li><li>• Power supply: 230-240V/50Hz - Vacuum capacity: 18 litres/mim - Maximum depression: -75kPa (-563mmHg)</li><li>• Vacuum is created by a plastic piston and cylinder system, with four vacuum-creating modules (QuatroFlex™ technology).</li><li>• The membrane vacuum regulator permits accurate vacuum settings.</li><li>• Double overflow-protection system (bottle and pump).</li><li>• Working temperature range: +5 to +40°C.</li></ul>
3.		LARYNGOSCOPE ADULT AND PEDIATRIC



	<b>Laryngoscope</b>	<ul style="list-style-type: none"> <li>• Should supply 4 different size standard blades and one handle for adult and pediatric separately and one short stubby handle .</li> <li>• Should be stainless Steel matt finished.</li> <li>• Should provide curved blades for both adult and pediatric.</li> <li>• An extra large blade should be supplied along with each scope.</li> <li>• Should be provided with battery 6 Should provide spare bulb – 6 nos</li> </ul>
4.	<b>Stilette for ET Tube</b>	
5.	<b>Meagles Forceps</b>	
6.	<b>Ambu Bags</b>	

**Terms & Conditions:**

1. The equipment should be of Good Quality
2. The equipment should meet standards in quality i.e ISO/TUV/CE/USFDA approved.
3. Supply should be **F.O.R** Destination and Installation.
4. Rates quoted should not be more than those quoted to **DGS&D** and any other Central or State Govts. Organizations
5. Taxes should be clearly mentioned.
6. Guarantee/Warranty should be quoted.

**Note: Only Terms & Conditions mentioned on this quotation will be considered for supply order.**

You are requested to send your lowest bid in sealed envelope, addressed to The PRINCIPAL, G.G.S Medical College, FARIDKOT super scribing " QUOTATION for "Equipments" on the top of Envelope.

Last Date for receipt of Quotation in Principal Office is **20.07.2015** by **5.00 Pm** through **Registered/Speed-Post Only**.

*Seave*  
Principal